საქართველოს ეროვნული სტანდარტი

წყლის ხარისხი. პოლიციკლური არომატული ჰიდროკარბონატის (PAH) განსაზღვრა ნაწილი 2. ექვსი PAH განსაზღვრა მაღალი გარჩევითობის მქონე თხევად ქრომატგრაფში ფლუორესცენტული მეთოდით. სითხე-სითხე ექსტრაქციის შემდეგ

საინფორმაციო მონაცემები

- 1 შემუშავებულია საქართველოს სტანდარტების, ტექნიკური რეგლამენტების და მეტროლოგიის ეროვნული სააგენტოს სტანდარტებისა და ტექნიკური რეგლამენტების დეპარტამენტის მიერ
- 2 დამტკიცებულია და შემოღებულია სამოქმედოდ საქართველოს სტანდარტების, ტექნიკური რეგლამენტების და მეტროლოგიის ეროვნული სააგენტოს 26 დეკემბრის #64 განკარგულებით
- 3 მიღებულია გარეკანის მეთოდით სტანდარტიზაციის საერთაშორისო ორგანიზაციის სტანდარტი ისო 7981-2 : 2005 "წყლის ხარისხი. პოლიციკლური არომატული ჰიდროკარბონატის (PAH) განსაზღვრა. ნაწილი 2. ექვსი PAH განსაზღვრა მაღალი გარჩევითობის მქონე თხევად ქრომატოგრაფში ფლუორესცენტული მეთოდით. სითხე-სითხე ექსტრაქციის შემდეგ"PA

4 პირველად

5 რეგისტრირებულია საქართველოს სტანდარტების, ტექნიკური რეგლამენტების და მეტროლოგიის ეროვნული სააგენტოს რეესტრში: 2007 წლის 26 დეკემბერი #268-1.3-0628

წინამდებარე სტანდარტის სრული ან ნაწილობრივი აღწარმოება, ტირაჟირება და გავრცელება საქართველოს სტანდარტების, ტექნიკური რეგლამენტების და მეტროლოგიის ეროვნული სააგენტოს ნებართვის გარეშე არ დაიშვება

INTERNATIONAL STANDARD

ISO 7981-2

First edition 2005-06-15

Water quality — Determination of polycyclic aromatic hydrocarbons (PAH) —

Part 2:

Determination of six PAH by highperformance liquid chromatography with fluorescence detection after liquid-liquid extraction

Qualité de l'eau — Détermination des hydrocarbures aromatiques polycycliques (HAP) —

Partie 2: Dosage de six HAP par chromatographie de haute performance en phase liquide avec détection fluorimétrique à la suite d'une extraction liquide-liquide



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

ForewordIntroduction					
			1	Scope	1
			2	Normative references	1
3	Principle	1			
4	Interferences	2			
5	Reagents	3			
6	Apparatus	4			
7	Sampling	5			
8	Procedure	6			
9	Calibration	8			
10	Measurement of samples	9			
11	Determination of the recovery	9			
12	Blank measurement	10			
13	Calculation	10			
14	Expression of results	10			
15	Test report	11			
16	Accuracy	11			
Anne	ex A (informative) Examples of chromatographic conditions and columns	12			
	ex B (informative) Examples for the construction of special apparatus				
	ex C (informative) Accuracy				
Bibli	ography	18			

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 7981-2 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

ISO 7981 consists of the following parts, under the general title *Water quality* — *Determination of polycyclic aromatic hydrocarbons (PAH)*:

- Part 1: Determination of six PAH by high-performance thin-layer chromatography with fluorescence detection after liquid-liquid extraction
- Part 2: Determination of six PAH by high-performance liquid chromatography with fluorescence detection after liquid-liquid extraction

Introduction

Polycyclic aromatic hydrocarbons (PAH) are present in nearly all types of waters. These substances are adsorbed on solids (sediments, suspended matter) as well as dissolved in the liquid phase.

Some PAH are known or suspected to cause cancer. The maximum acceptable levels of PAH in waters intended for human consumption are given in European Legislation [1] [2] [3] [4].

The sum of the mass concentrations of the six PAH specified in this part of ISO 7981 usually is about 0,01 μ g/l to 0,05 μ g/l in ground water, up to 1 μ g/l in surface water, and up to 1 000 μ g/l in waste water.